

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062520\
 Data File : VU039116.D
 Acq On : 25 Jun 2020 15:15
 Operator : SY/MD
 Sample : L3109-13DL 2X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9F57DL

Quant Time: Jun 26 12:26:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 26 00:59:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	254452	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	256475	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	117584	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	86831	4.66	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.20%
7) Chloroethane-d5	1.93	69	80047	5.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.00%
11) 1,1-Dichloroethene-d2	2.58	63	133325	3.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.40%
20) 2-Butanone-d5	4.68	46	420381	62.51	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.02%
24) Chloroform-d	5.10	84	197220	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.60%
26) 1,2-Dichloroethane-d4	5.74	65	125304	5.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.20%
32) Benzene-d6	5.76	84	373170	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.00%
36) 1,2-Dichloropropane-d6	6.72	67	126937	5.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.20%
41) Toluene-d8	7.92	98	285129	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.80%
43) trans-1,3-Dichloropropene-	8.20	79	51857	4.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.80%
46) 2-Hexanone-d5	8.66	63	332220	61.97	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.94%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	118402	5.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.60%
64) 1,2-Dichlorobenzene-d4	12.20	152	120756	5.51	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.20%


Target Compounds					Ovalue
34) Trichloroethene	6.57	95	11170	0.573 ug/L	94
47) Tetrachloroethene	8.57	164	140132	9.975 ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062520\
Data File : VU039116.D
Acq On : 25 Jun 2020 15:15
Operator : SY/MD
Sample : L3109-13DL 2X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9F57DL

Quant Time: Jun 26 12:26:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 26 00:59:08 2020
Response via : Initial Calibration









#34
 Trichloroethene
 Concen: 0.573 ug/L
 RT: 6.57 min Scan# 1626
 Delta R.T. -0.00 min
 Lab File: VU039116.D
 Acq: 25 Jun 2020 15:15

Instrument :
 MSVOA_U
 ClientSampleId :
 F9F57DL

Tot Ion:	95	Resp:	11170
Ion	Ratio	Lower	Upper
95	100		
97	62.3	45.1	83.7
132	78.6	65.1	120.9
130	98.1	68.6	127.4



#47
 Tetrachloroethene
 Concen: 9.975 ug/L
 RT: 8.57 min Scan# 2249
 Delta R.T. -0.00 min
 Lab File: VU039116.D
 Acq: 25 Jun 2020 15:15

Tgt Ion:	164	Resp:	140132
Ion	Ratio	Lower	Upper
164	100		
129	96.6	71.3	132.5
131	94.7	68.3	126.8
166	127.2	90.8	168.6

