

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062520\
 Data File : VU039130.D
 Acq On : 25 Jun 2020 22:23
 Operator : SY/MD
 Sample : L3109-21
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jun 26 04:34:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 26 04:20:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	241510	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.43	117	241191	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	102752	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	68880	3.90	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.00%
7) Chloroethane-d5	1.93	69	66989	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.60%
11) 1,1-Dichloroethene-d2	2.59	63	104237	3.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.40%
20) 2-Butanone-d5	4.67	46	348760	54.64	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.28%
24) Chloroform-d	5.10	84	164778	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
26) 1,2-Dichloroethane-d4	5.74	65	101677	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
32) Benzene-d6	5.76	84	294851	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
36) 1,2-Dichloropropane-d6	6.72	67	104691	4.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.20%
41) Toluene-d8	7.92	98	222491	3.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.60%
43) trans-1,3-Dichloropropene-	8.20	79	41795	4.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.60%
46) 2-Hexanone-d5	8.65	63	252057	50.00	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.00%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	95357	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.20%
64) 1,2-Dichlorobenzene-d4	12.20	152	94123	4.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.40%


Target Compounds					Ovalue
16) Methylene chloride	3.08	84	3862	0.187 ug/L	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062520\
Data File : VU039130.D
Acq On : 25 Jun 2020 22:23
Operator : SY/MD
Sample : L3109-21
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jun 26 04:34:31 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 26 04:20:15 2020
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.187 ug/L
RT: 3.08 min Scan# 541
Delta R.T. 0.00 min
Lab File: VU039130.D
Acq: 25 Jun 2020 22:23

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Tot Ion: 84 Resp: 3862

Ion	Ratio	Lower	Upper
84	100		
86	47.7	44.1	81.9
49	119.1	93.1	172.9

