

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062618\
 Data File : VU024924.D
 Acq On : 26 Jun 2018 14:01
 Operator : MD/SY
 Sample : J3621-07ME
 Misc : 7.46g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DAZ59ME

Quant Time: Jun 27 01:22:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 27 01:09:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	369843	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.10	117	336851	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	161262	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	67098	31.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	63.40%
7) Chloroethane-d5	1.66	69	40632	24.95	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	49.90%#
11) 1,1-Dichloroethene-d2	2.24	63	120892	28.58	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.16%#
21) 2-Butanone-d5	4.20	46	167495	93.74	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.74%
24) Chloroform-d	4.65	84	182978	42.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.80%
26) 1,2-Dichloroethane-d4	5.31	65	132651	49.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.18%
32) Benzene-d6	5.34	84	353653	43.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.24%
36) 1,2-Dichloropropane-d6	6.33	67	119206	43.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.62%
41) Toluene-d8	7.57	98	346413	41.86	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.72%
43) trans-1,3-Dichloropropene-	7.85	79	57536	41.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.88%
47) 2-Hexanone-d5	8.34	63	133308	95.17	ug/L	0.03
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.17%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	182757	45.79	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.58%
64) 1,2-Dichlorobenzene-d4	11.87	152	147098	47.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.54%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062618\
Data File : VU024924.D
Acq On : 26 Jun 2018 14:01
Operator : MD/SY
Sample : J3621-07ME
Misc : 7.46g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DAZ59ME

Quant Time: Jun 27 01:22:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061518WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 27 01:09:35 2018
Response via : Initial Calibration

