

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062620\
 Data File : VU039157.D
 Acq On : 26 Jun 2020 11:52
 Operator : SY/MD
 Sample : L3121-01DL 5X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9F00DL

Quant Time: Jun 27 00:20:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 26 23:52:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	272593	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	280253	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	121216	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	52537	2.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	52.60%
7) Chloroethane-d5	1.93	69	53353	3.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	64.60%#
11) 1,1-Dichloroethene-d2	2.59	63	84134	2.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	43.80%#
20) 2-Butanone-d5	4.68	46	304558	42.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.56%
24) Chloroform-d	5.10	84	137297	3.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	70.00%
26) 1,2-Dichloroethane-d4	5.74	65	88162	3.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.00%
32) Benzene-d6	5.76	84	249879	3.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	64.40%#
36) 1,2-Dichloropropane-d6	6.72	67	87471	3.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	67.00%
41) Toluene-d8	7.92	98	191457	2.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	59.00%#
43) trans-1,3-Dichloropropene-	8.20	79	35735	3.15	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	63.00%
46) 2-Hexanone-d5	8.66	63	216343	36.93	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.86%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	83525	3.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	68.80%
64) 1,2-Dichlorobenzene-d4	12.20	152	81928	3.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	72.60%#

Target Compounds					Ovalue
34) Trichloroethene	6.57	95	9972	0.468 ug/L	91
47) Tetrachloroethene	8.57	164	97594	6.358 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062620\
Data File : VU039157.D
Acq On : 26 Jun 2020 11:52
Operator : SY/MD
Sample : L3121-01DL 5X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9F00DL

Quant Time: Jun 27 00:20:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 26 23:52:58 2020
Response via : Initial Calibration

