

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024937.D
 Acq On : 26 Jun 2018 22:20
 Operator : MD/SY
 Sample : J3669-16ME 20X
 Misc : 5.94g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB483-35-37-180620ME

Manual Integrations
 APPROVED

apatel
 6/28/2018 4:38:18 PM

Quant Time: Jun 27 04:51:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	206779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	314385	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	288913	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	156072	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	156210	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
35) Dibromofluoromethane	4.89	113	120708	46.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.52%	
50) Toluene-d8	7.57	98	449182	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
62) 4-Bromofluorobenzene	10.31	95	167079	44.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.44%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	4488m	1.51	ug/l	Qvalue
44) Trichloroethene	6.19	130	245645	81.99	ug/l	97
52) Toluene	7.64	92	6607	0.98	ug/l	100
91) 1,2-Dichlorobenzene	11.89	146	4037	0.72	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024937.D
Acq On : 26 Jun 2018 22:20
Operator : MD/SY
Sample : J3669-16ME 20X
Misc : 5.94g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB483-35-37-180620ME

Manual Integrations
APPROVED

apatel
6/28/2018 4:38:18 PM

Quant Time: Jun 27 04:51:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

