

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024941.D
 Acq On : 26 Jun 2018 23:57
 Operator : MD/SY
 Sample : J3669-30ME 5X
 Misc : 6.29g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB474-31-33-180620ME

Quant Time: Jun 27 09:29:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	211371	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	321226	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	295647	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	164966	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	157598	45.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.34%	
35) Dibromofluoromethane	4.89	113	123588	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	
50) Toluene-d8	7.57	98	462368	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
62) 4-Bromofluorobenzene	10.31	95	174041	45.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.16%	
Target Compounds						
4) Vinyl Chloride	1.40	62	3533	1.27	ug/l	97
12) 1,1-Dichloroethene	2.28	96	4070	1.69	ug/l	94
27) cis-1,2-Dichloroethene	4.23	96	144537	47.43	ug/l	88
44) Trichloroethene	6.19	130	190777	62.32	ug/l	94
65) Chlorobenzene	9.13	112	4128	0.55	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024941.D
Acq On : 26 Jun 2018 23:57
Operator : MD/SY
Sample : J3669-30ME 5X
Misc : 6.29g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB474-31-33-180620ME

Quant Time: Jun 27 09:29:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

