

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024944.D
 Acq On : 27 Jun 2018 01:10
 Operator : MD/SY
 Sample : J3669-25ME 10X
 Misc : 5.96g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB479-31-33-180620ME

Quant Time: Jun 27 05:31:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	213723	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	323118	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	297817	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	157171	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	161640	46.33	ug/l	0.00
Spiked Amount			Recovery	=	92.66%	
35) Dibromofluoromethane	4.89	113	123666	46.12	ug/l	0.00
Spiked Amount			Recovery	=	92.24%	
50) Toluene-d8	7.57	98	467100	47.75	ug/l	0.00
Spiked Amount			Recovery	=	95.50%	
62) 4-Bromofluorobenzene	10.31	95	172309	44.37	ug/l	0.00
Spiked Amount			Recovery	=	88.74%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
27) cis-1,2-Dichloroethene	4.23	96	37550	12.19	ug/l	90
44) Trichloroethene	6.19	130	115996	37.67	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024944.D
Acq On : 27 Jun 2018 01:10
Operator : MD/SY
Sample : J3669-25ME 10X
Misc : 5.96a/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB479-31-33-180620ME

Quant Time: Jun 27 05:31:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

