

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024947.D
 Acq On : 27 Jun 2018 02:22
 Operator : MD/SY
 Sample : J3664-23ME 20X
 Misc : 6.77g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB471-33-35-180619ME

Quant Time: Jun 27 05:41:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	208933	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	317360	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	296231	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	156739	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	159588	46.79	ug/l	0.00
Spiked Amount			Recovery	=	93.58%	
35) Dibromofluoromethane	4.89	113	122798	46.62	ug/l	0.00
Spiked Amount			Recovery	=	93.24%	
50) Toluene-d8	7.57	98	460314	47.91	ug/l	0.00
Spiked Amount			Recovery	=	95.82%	
62) 4-Bromofluorobenzene	10.31	95	169417	44.41	ug/l	0.00
Spiked Amount			Recovery	=	88.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
12) 1,1-Dichloroethene	2.29	96	2808	1.18	ug/l	88
20) Methylene Chloride	2.70	84	3889	1.35	ug/l	93
27) cis-1,2-Dichloroethene	4.23	96	1677	0.56	ug/l	89
44) Trichloroethene	6.19	130	421579	139.40	ug/l	96
52) Toluene	7.64	92	13832	2.02	ug/l	98
91) 1,2-Dichlorobenzene	11.89	146	5350	0.95	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024947.D
Acq On : 27 Jun 2018 02:22
Operator : MD/SY
Sample : J3664-23ME 20X
Misc : 6.77g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB471-33-35-180619ME

Quant Time: Jun 27 05:41:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

