

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024951.D
 Acq On : 27 Jun 2018 03:59
 Operator : MD/SY
 Sample : J3664-19ME
 Misc : 4.59g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-0485-180619ME

Manual Integrations
 APPROVED

MMDadoda
 6/28/2018 4:37:25 PM

Quant Time: Jun 27 09:47:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	213938	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	329739	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.10	117	319053	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	201766	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	166816	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
35) Dibromofluoromethane	4.89	113	125167	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
50) Toluene-d8	7.57	98	479946	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
62) 4-Bromofluorobenzene	10.32	95	209830	52.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.88%	
Target Compounds						
39) Methylcyclohexane	6.41	83	7432	1.54	ug/l	# 88
65) Chlorobenzene	9.13	112	30092	3.69	ug/l	94
83) tert-Butylbenzene	11.11	119	13622	1.05	ug/l	79
85) sec-Butylbenzene	11.33	105	179154	11.08	ug/l	99
89) n-Butylbenzene	11.90	91	185030m	14.39	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024951.D
Acq On : 27 Jun 2018 03:59
Operator : MD/SY
Sample : J3664-19ME
Misc : 4.59g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-0485-180619ME

Manual Integrations
APPROVED

MMDadoda
6/28/2018 4:37:25 PM

Quant Time: Jun 27 09:47:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

