

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024952.D
 Acq On : 27 Jun 2018 04:23
 Operator : MD/SY
 Sample : J3664-13ME
 Misc : 5.82g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB485-21-23-180619ME

Manual Integrations
 APPROVED

MMDadoda
 6/28/2018 4:37:26 PM

Quant Time: Jun 27 06:33:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	242948	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	363320	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.10	117	350396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	230409	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	179667	45.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.60%	
35) Dibromofluoromethane	4.89	113	138882	46.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.12%	
50) Toluene-d8	7.57	98	532234	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
62) 4-Bromofluorobenzene	10.32	95	242201	55.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.92%	
Target Compounds						
39) Methylcyclohexane	6.41	83	14801	2.78	ug/l	96
65) Chlorobenzene	9.13	112	83864	9.37	ug/l	98
73) Isopropylbenzene	10.17	105	12539	0.71	ug/l	98
83) tert-Butylbenzene	11.11	119	30843	2.09	ug/l	97
85) sec-Butylbenzene	11.33	105	473891	25.67	ug/l	98
89) n-Butylbenzene	11.90	91	482041m	32.84	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024952.D
Acq On : 27 Jun 2018 04:23
Operator : MD/SY
Sample : J3664-13ME
Misc : 5.82g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB485-21-23-180619ME

Manual Integrations
APPROVED

MMDadoda
6/28/2018 4:37:26 PM

Quant Time: Jun 27 06:33:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

