

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024963.D
 Acq On : 27 Jun 2018 13:50
 Operator : MD/SY
 Sample : J3669-19MEDL 5X
 Misc : 6.02g/5.0mL/100uL/5.00ML/MSVOA_U/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB475-27-29-180620MEDL

Quant Time: Jun 28 02:54:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	229651	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	340685	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	312212	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	172198	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	170424	45.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.92%	
35) Dibromofluoromethane	4.89	113	132576	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
50) Toluene-d8	7.57	98	484086	46.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.88%	
62) 4-Bromofluorobenzene	10.31	95	183287	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
Target Compounds						
4) Vinyl Chloride	1.40	62	7512	2.48	ug/l	96
27) cis-1,2-Dichloroethene	4.23	96	92330	27.89	ug/l	87
44) Trichloroethene	6.19	130	42146	12.98	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024963.D
Acq On : 27 Jun 2018 13:50
Operator : MD/SY
Sample : J3669-19MEDL 5X
Misc : 6.02a/5.0mL/100uL/5.00ML/MSVOA_U/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB475-27-29-180620MEDL

Quant Time: Jun 28 02:54:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

