

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024966.D
 Acq On : 27 Jun 2018 15:03
 Operator : MD/SY
 Sample : J3604-02DL 4X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-10DL

Quant Time: Jun 28 12:21:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	229428	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	344980	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	315144	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	170525	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	171341	45.75	ug/l	0.00
Spiked Amount			Recovery	=	91.50%	
35) Dibromofluoromethane	4.89	113	134018	46.81	ug/l	0.00
Spiked Amount			Recovery	=	93.62%	
50) Toluene-d8	7.57	98	491457	47.06	ug/l	0.00
Spiked Amount			Recovery	=	94.12%	
62) 4-Bromofluorobenzene	10.31	95	184529	44.50	ug/l	0.00
Spiked Amount			Recovery	=	89.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
64) Tetrachloroethene	8.23	164	122825	41.00	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024966.D
Acq On : 27 Jun 2018 15:03
Operator : MD/SY
Sample : J3604-02DL 4X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-10DL

Quant Time: Jun 28 12:21:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

