

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024969.D
 Acq On : 27 Jun 2018 16:15
 Operator : MD/SY
 Sample : J3664-12ME
 Misc : 5.61g/5.0mL/100uL/5.00ML/MSVOA_U/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-0476-180619ME

Quant Time: Jun 28 04:15:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	242634	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	368448	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	350756	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	194293	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	182134	45.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.96%	
35) Dibromofluoromethane	4.89	113	139592	45.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.30%	
50) Toluene-d8	7.57	98	531134	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
62) 4-Bromofluorobenzene	10.32	95	207630	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
21) trans-1,2-Dichloroethene	2.97	96	3273	1.08	ug/l	92
27) cis-1,2-Dichloroethene	4.22	96	25194	7.20	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024969.D
Acq On : 27 Jun 2018 16:15
Operator : MD/SY
Sample : J3664-12ME
Misc : 5.61g/5.0mL/100uL/5.00ML/MSVOA_U/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-0476-180619ME

Quant Time: Jun 28 04:15:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

