

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024971.D
 Acq On : 27 Jun 2018 17:04
 Operator : MD/SY
 Sample : J3669-20ME 20X
 Misc : 6.20g/5.0mL/100uL/5.00ML/MSVOA_U/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB475-29-31-180620ME

Quant Time: Jun 28 04:38:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	227189	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	335634	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	310178	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	167653	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	169908	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
35) Dibromofluoromethane	4.89	113	129698	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
50) Toluene-d8	7.57	98	484477	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
62) 4-Bromofluorobenzene	10.31	95	181697	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
27) cis-1,2-Dichloroethene	4.23	96	3194	0.98	ug/l	95
44) Trichloroethene	6.19	130	3070	0.96	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024971.D
Acq On : 27 Jun 2018 17:04
Operator : MD/SY
Sample : J3669-20ME 20X
Misc : 6.20g/5.0mL/100uL/5.00ML/MSVOA_U/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB475-29-31-180620ME

Quant Time: Jun 28 04:38:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

