

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024977.D
 Acq On : 27 Jun 2018 19:42
 Operator : MD/SY
 Sample : J3670-05ME
 Misc : 6.00µ/5.0mL/100µL/5.00ML/MSVOA_U/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB487-17-19-180621ME

Quant Time: Jun 28 04:53:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	229220	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	350553	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.10	117	334951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	192948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	174083	46.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.04%	
35) Dibromofluoromethane	4.89	113	135458	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
50) Toluene-d8	7.57	98	507900	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
62) 4-Bromofluorobenzene	10.32	95	208287	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.22	96	5488	1.66	ug/l	98
65) Chlorobenzene	9.13	112	39070	4.57	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024977.D
Acq On : 27 Jun 2018 19:42
Operator : MD/SY
Sample : J3670-05ME
Misc : 6.00µL/5.0mL/100µL/5.00ML/MSVOA_U/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB487-17-19-180621ME

Quant Time: Jun 28 04:53:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

