

Data File : VU024985.D
Acq On : 27 Jun 2018 23:20
Operator : MD/SY
Sample : J3670-06ME
Misc : 5.79g/5mL/100uL/5.00ML/MSVOA_U/MEOH
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB487-19-21-180621ME

Manual Integrations
APPROVED

MMDadoda
6/28/2018 4:37:06 PM

Quant Time: Jul 06 05:02:15 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	235824	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	358139	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.10	117	340735	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	191149	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	176145	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	
35) Dibromofluoromethane	4.89	113	136966	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
50) Toluene-d8	7.57	98	517552	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
62) 4-Bromofluorobenzene	10.32	95	211810	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.21	96	5318m	1.564	ug/l	
65) Chlorobenzene	9.12	112	19453	2.235	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU024985.D
Acq On : 27 Jun 2018 23:20
Operator : MD/SY
Sample : J3670-06ME
Misc : 5.79g/5mL/100uL/5.00ML/MSVOA_U/MEOH
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB487-19-21-180621ME

Manual Integrations
APPROVED

MMDadoda
6/28/2018 4:37:06 PM

Quant Time: Jul 06 05:02:15 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

