

Data File : VU024996.D
Acq On : 28 Jun 2018 03:46
Operator : MD/SY
Sample : J3671-11ME 10X
Misc : 6.26g/5mL/100uL/5.00ML/MSVOA_U/MEOH
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB478-29-31-180621ME

Quant Time: Jul 05 03:24:05 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	213014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	325694	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	302290	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	165751	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	163087	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
35) Dibromofluoromethane	4.89	113	125566	46.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.90%	
50) Toluene-d8	7.57	98	466644	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	
62) 4-Bromofluorobenzene	10.31	95	179425	45.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.66%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	152171	49.547	ug/l	89
44) Trichloroethene	6.19	130	94972	30.600	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU024996.D
Acq On : 28 Jun 2018 03:46
Operator : MD/SY
Sample : J3671-11ME 10X
Misc : 6.26g/5mL/100uL/5.00ML/MSVOA_U/MEOH
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB478-29-31-180621ME

Quant Time: Jul 05 03:24:05 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

