

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062718\
 Data File : VU024998.D
 Acq On : 28 Jun 2018 04:34
 Operator : MD/SY
 Sample : J3671-08ME 10X
 Misc : 6.43g/5mL/100uL/5.00ML/MSVOA_U/MEOH
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB478-21-23-180621ME

Quant Time: Jun 28 07:06:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	214710	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	324865	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	306841	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	169239	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	164319	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
35) Dibromofluoromethane	4.89	113	125651	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
50) Toluene-d8	7.57	98	469372	47.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.46%	
62) 4-Bromofluorobenzene	10.31	95	178593	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	3864	1.36	ug/l	99
27) cis-1,2-Dichloroethene	4.23	96	157163	50.77	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062718\
Data File : VU024998.D
Acq On : 28 Jun 2018 04:34
Operator : MD/SY
Sample : J3671-08ME 10X
Misc : 6.43g/5mL/100uL/5.00ML/MSVOA_U/MEOH
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB478-21-23-180621ME

Quant Time: Jun 28 07:06:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 4.99 min Scan# 1367

Delta R.T. -0.00 min

Lab File: VU024998.D

Acq: 28 Jun 2018 04:34

Instrument :

MSVOA_U

ClientSampleId :


WP021SB478-21-23-180621ME

Tot Ion:168 Resp: 214710

Ion Ratio Lower Upper

168 100

99 55.9 43.2 64.8



Abundance Ion 167.90 (167.60 to 168.60): VU024502.D (-1351) (-)

Ion 98.90 (98.60 to 99.60): VU024502.D (-1351) (-)

4.99

100000

80000



60000

40000

20000

0

Time-->



#4

Vinyl Chloride

Concen: 1.36 ug/l

RT: 1.40 min Scan# 253

Delta R.T. -0.00 min

Lab File: VU024998.D


Acq: 28 Jun 2018 04:34

Tgt Ion: 62 Resp: 3864

Ion Ratio Lower Upper

62 100

64 31.8 24.8 37.2



Abundance Ion 61.90 (61.60 to 62.60): VU024502.D (-244) (-)

Ion 63.90 (63.60 to 64.60): VU024502.D (-244) (-)

1.40

4000


3000


2000

1000

0

Time-->





#27

cis-1,2-Dichloroethene

Concen: 50.77 ug/l

RT: 4.23 min Scan# 1132

Delta R.T. -0.00 min

Lab File: VU024998.D

Acq: 28 Jun 2018 04:34

Instrument :

MSVOA_U

ClientSampleId :

WP021SB478-21-23-180621ME





Tot Ion: 96 Resp: 157163

Ion Ratio Lower Upper

96 100

61 132.8 0.0 306.6

98 63.5 0.0 128.8



#33

1,2-Dichloroethane-d4

Concen: 46.88 ug/l

RT: 5.31 min Scan# 1469

Delta R.T. -0.00 min

Lab File: VU024998.D




Acq: 28 Jun 2018 04:34


Tgt Ion: 65 Resp: 164319

Ion Ratio Lower Upper

65 100

67 50.3 0.0 107.0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 5.89 min Scan# 1648

Delta R.T. -0.00 min

Lab File: VU024998.D

Acq: 28 Jun 2018 04:34

Instrument :

MSVOA_U

ClientSampleId :

WP021SB478-21-23-180621ME


Tot Ion:114 Resp: 324865

Ion Ratio Lower Upper

114 100

63 21.1 0.0 45.4




88 15.7 0.0 31.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VU0

Ion 87.90 (87.60 to 88.60): VU0



#35

Dibromofluoromethane

Concen: 46.60 ug/l

RT: 4.89 min Scan# 1337

Delta R.T. 0.00 min

Lab File: VU024998.D

Acq: 28 Jun 2018 04:34


Tgt Ion:113 Resp: 125651

Ion Ratio Lower Upper

113 100

111 102.5 82.2 123.4



192 21.7 16.2 24.4



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 47.73 ug/l

RT: 7.57 min Scan# 2170

Delta R.T. -0.00 min

Lab File: VU024998.D

Acq: 28 Jun 2018 04:34

Instrument :

MSVOA_U

ClientSampleId :





WP021SB478-21-23-180621ME

Tot Ion: 98 Resp: 469372

Ion Ratio Lower Upper

98 100

100 64.0 51.1 76.7



#62

4-Bromofluorobenzene

Concen: 45.74 ug/l

RT: 10.31 min Scan# 3023

Delta R.T. -0.00 min

Lab File: VU024998.D

Acq: 28 Jun 2018 04:34




Tgt Ion: 95 Resp: 178593


Ion Ratio Lower Upper

95 100

174 85.0 0.0 165.8

176 82.8 0.0 159.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 9.09 min Scan# 2645
Delta R.T. -0.00 min
Lab File: VU024998.D
Acq: 28 Jun 2018 04:34

Instrument :
MSVOA_U
ClientSampleId :
WP021SB478-21-23-180621ME

Tot Ion	Ratio	Lower	Upper
117	100		
82	55.1	44.3	66.5
119	32.2	25.4	38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 11.49 min Scan# 3389
Delta R.T. -0.00 min
Lab File: VU024998.D
Acq: 28 Jun 2018 04:34

Tgt Ion	Ratio	Lower	Upper
152	100		
115	57.3	43.0	129.0
150	154.4	0.0	354.0

