

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049405.D
 Acq On : 27 Jun 2022 15:22
 Operator : SY/MD
 Sample : VU0627MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0627MBL01

Quant Time: Jun 27 19:05:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 27 14:37:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	159904	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.246	114	264351	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	219878	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	89273	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	114544	46.513	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.020%
35) Dibromofluoromethane	5.288	113	87246	47.174	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.340%
50) Toluene-d8	7.896	98	319322	48.407	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.820%
62) 4-Bromofluorobenzene	10.635	95	101174	40.038	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.080%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049405.D
Acq On : 27 Jun 2022 15:22
Operator : SY/MD
Sample : VU0627MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0627MBL01

Quant Time: Jun 27 19:05:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 27 14:37:12 2022
Response via : Initial Calibration

