

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049406.D
 Acq On : 27 Jun 2022 16:14
 Operator : SY/MD
 Sample : VU0627WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0627WBL01

Quant Time: Jun 27 19:05:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 27 14:37:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	182859	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	317770	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	283985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	124530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	136999	48.648	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.300%
35) Dibromofluoromethane	5.292	113	104393	46.956	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.920%
50) Toluene-d8	7.899	98	384879	48.537	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.080%
62) 4-Bromofluorobenzene	10.632	95	132433	43.599	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049406.D
Acq On : 27 Jun 2022 16:14
Operator : SY/MD
Sample : VU0627WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0627WBL01

Quant Time: Jun 27 19:05:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 27 14:37:12 2022
Response via : Initial Calibration

