

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049411.D
 Acq On : 27 Jun 2022 18:17
 Operator : SY/MD
 Sample : N3431-23
 Misc : 3.12g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-14B-2-5-GR

Quant Time: Jun 27 19:08:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 27 14:37:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	266159	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	468675	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	426813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	207764	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	183565	44.783	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.560%
35) Dibromofluoromethane	5.289	113	140982	42.996	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.000%
50) Toluene-d8	7.896	98	588115	50.287	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.580%
62) 4-Bromofluorobenzene	10.639	95	245560	54.812	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.620%
Target Compounds						
					Qvalue	
39) Methylcyclohexane	6.729	83	105765	18.559	ug/l	# 13
40) Benzene	5.765	78	233746	16.284	ug/l	99
52) Toluene	7.967	92	68960	8.083	ug/l	96
67) Ethyl Benzene	9.572	91	72752	4.542	ug/l	95
68) m/p-Xylenes	9.697	106	97646	16.475	ug/l	88
69) o-Xylene	10.105	106	8353	1.420	ug/l	100
73) Isopropylbenzene	10.488	105	180209	12.176	ug/l	96
78) n-propylbenzene	10.909	91	118475	6.588	ug/l	94
80) 1,3,5-Trimethylbenzene	11.092	105	8276	0.641	ug/l	97
84) 1,2,4-Trimethylbenzene	11.472	105	66551	5.243	ug/l	93
85) sec-Butylbenzene	11.645	105	233172	14.593	ug/l	95
89) n-Butylbenzene	12.211	91	88021	7.359	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049411.D
Acq On : 27 Jun 2022 18:17
Operator : SY/MD
Sample : N3431-23
Misc : 3.12g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-14B-2-5-GR

Quant Time: Jun 27 19:08:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 27 14:37:12 2022
Response via : Initial Calibration

