

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049420.D
 Acq On : 27 Jun 2022 22:45
 Operator : SY/MD
 Sample : VU0627MBL02
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0627MBL02

Quant Time: Jun 28 05:29:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 05:14:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	295686	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	502361	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	447442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	207041	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	190460	41.825	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.640%
35) Dibromofluoromethane	5.288	113	154528	43.967	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.940%
50) Toluene-d8	7.896	98	598045	47.707	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.420%
62) 4-Bromofluorobenzene	10.632	95	218319	45.464	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049420.D
Acq On : 27 Jun 2022 22:45
Operator : SY/MD
Sample : VU0627MBL02
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0627MBL02

Quant Time: Jun 28 05:29:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 05:14:22 2022
Response via : Initial Calibration

