

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049421.D
 Acq On : 27 Jun 2022 23:08
 Operator : SY/MD
 Sample : VU0627WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0627WBL02

Quant Time: Jun 28 05:30:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 05:14:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	300802	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	516876	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	457468	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	209108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	197376	42.606	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.220%
35) Dibromofluoromethane	5.288	113	159861	44.207	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.420%
50) Toluene-d8	7.896	98	612532	47.490	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.980%
62) 4-Bromofluorobenzene	10.632	95	219625	44.451	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.900%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049421.D
Acq On : 27 Jun 2022 23:08
Operator : SY/MD
Sample : VU0627WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0627WBL02

Quant Time: Jun 28 05:30:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 05:14:22 2022
Response via : Initial Calibration

