

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049424.D
 Acq On : 28 Jun 2022 00:19
 Operator : SY/MD
 Sample : N3386-16DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE134D3-20220616DL

Quant Time: Jun 28 05:32:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 05:14:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	315589	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	531108	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	474325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	220617	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	202172	41.597	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.200%
35) Dibromofluoromethane	5.288	113	163951	44.123	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.240%
50) Toluene-d8	7.896	98	634129	47.847	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.700%
62) 4-Bromofluorobenzene	10.632	95	230665	45.435	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.860%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.578	101	105006	29.993	ug/l	95
12) 1,1-Dichloroethene	2.578	96	942	0.275	ug/l	83
27) cis-1,2-Dichloroethene	4.668	96	1727	0.405	ug/l	62
44) Trichloroethene	6.542	130	52629	13.610	ug/l	88
64) Tetrachloroethene	8.555	164	30323	9.967	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049424.D
Acq On : 28 Jun 2022 00:19
Operator : SY/MD
Sample : N3386-16DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE134D3-20220616DL

Quant Time: Jun 28 05:32:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 05:14:22 2022
Response via : Initial Calibration

