

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049427.D
 Acq On : 28 Jun 2022 01:30
 Operator : SY/MD
 Sample : N3386-01RE 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE105D2-20220613RE

Quant Time: Jun 28 07:12:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 05:14:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	308730	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	529977	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	468899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	217513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	204637	43.039	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.080%
35) Dibromofluoromethane	5.288	113	162409	43.801	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.600%
50) Toluene-d8	7.896	98	630341	47.663	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.320%
62) 4-Bromofluorobenzene	10.632	95	226417	44.693	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.380%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.578	101	2002	0.585	ug/l	92
44) Trichloroethene	6.542	130	214320	55.542	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049427.D
Acq On : 28 Jun 2022 01:30
Operator : SY/MD
Sample : N3386-01RE 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE105D2-20220613RE

Quant Time: Jun 28 07:12:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 05:14:22 2022
Response via : Initial Calibration

