

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049431.D
 Acq On : 28 Jun 2022 03:04
 Operator : SY/MD
 Sample : N3386-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BPOW1-7-20220614

Quant Time: Jun 28 07:14:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 05:14:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	310848	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	537527	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	479245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	222090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	205720	42.972	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.940%
35) Dibromofluoromethane	5.288	113	165072	43.894	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.780%
50) Toluene-d8	7.896	98	635132	47.351	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.700%
62) 4-Bromofluorobenzene	10.632	95	231042	44.965	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.940%
Target Compounds						
16) Acetone	2.636	43	3078	1.275	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049431.D
Acq On : 28 Jun 2022 03:04
Operator : SY/MD
Sample : N3386-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BPOW1-7-20220614

Quant Time: Jun 28 07:14:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 05:14:22 2022
Response via : Initial Calibration

