

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049432.D
 Acq On : 28 Jun 2022 03:28
 Operator : SY/MD
 Sample : N3386-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE108D1-20220614

Quant Time: Jun 28 07:14:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 05:14:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	312635	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	534489	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	479006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	223991	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	204136	42.398	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.800%
35) Dibromofluoromethane	5.288	113	166972	44.652	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.300%
50) Toluene-d8	7.896	98	635612	47.656	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.320%
62) 4-Bromofluorobenzene	10.632	95	234911	45.978	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.960%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.575	101	909	0.262	ug/l	# 84
44) Trichloroethene	6.542	130	85829	22.055	ug/l	90
64) Tetrachloroethene	8.552	164	3933	1.280	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049432.D
Acq On : 28 Jun 2022 03:28
Operator : SY/MD
Sample : N3386-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE108D1-20220614

Quant Time: Jun 28 07:14:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 05:14:22 2022
Response via : Initial Calibration

