

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU062722\  
 Data File : VU049436.D  
 Acq On : 28 Jun 2022 05:02  
 Operator : SY/MD  
 Sample : N3482-08  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 43 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 FB062222

Quant Time: Jun 28 07:15:36 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U062722W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jun 28 05:14:22 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)     |
|-----------------------------|--------|-------|----------|----------|-------|--------------|
| -----                       |        |       |          |          |       |              |
| Internal Standards          |        |       |          |          |       |              |
| 1) Pentafluorobenzene       | 5.372  | 168   | 307699   | 50.000   | ug/l  | 0.00         |
| 34) 1,4-Difluorobenzene     | 6.250  | 114   | 530372   | 50.000   | ug/l  | 0.00         |
| 63) Chlorobenzene-d5        | 9.417  | 117   | 475429   | 50.000   | ug/l  | 0.00         |
| 72) 1,4-Dichlorobenzene-d4  | 11.812 | 152   | 220361   | 50.000   | ug/l  | 0.00         |
| System Monitoring Compounds |        |       |          |          |       |              |
| 33) 1,2-Dichloroethane-d4   | 5.703  | 65    | 202513   | 42.735   | ug/l  | 0.00         |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 85.480%      |
| 35) Dibromofluoromethane    | 5.288  | 113   | 165030   | 44.475   | ug/l  | 0.00         |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 88.960%      |
| 50) Toluene-d8              | 7.896  | 98    | 633434   | 47.861   | ug/l  | 0.00         |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 95.720%      |
| 62) 4-Bromofluorobenzene    | 10.632 | 95    | 231791   | 45.720   | ug/l  | 0.00         |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 91.440%      |
| Target Compounds            |        |       |          |          |       |              |
| 16) Acetone                 | 2.639  | 43    | 3603     | 1.508    | ug/l  | Qvalue<br>90 |
| -----                       |        |       |          |          |       |              |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU062722\  
Data File : VU049436.D  
Acq On : 28 Jun 2022 05:02  
Operator : SY/MD  
Sample : N3482-08  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 43 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
FB062222

Quant Time: Jun 28 07:15:36 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U062722W.M  
Quant Title : SW846 8260  
QLast Update : Tue Jun 28 05:14:22 2022  
Response via : Initial Calibration

