

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049440.D
 Acq On : 28 Jun 2022 06:36
 Operator : SY/MD
 Sample : N3482-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-38

Quant Time: Jun 28 07:16:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 05:14:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	299576	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	515180	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	461121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	212129	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	201956	43.773	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.540%
35) Dibromofluoromethane	5.292	113	161692	44.861	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.720%
50) Toluene-d8	7.899	98	609826	47.436	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.880%
62) 4-Bromofluorobenzene	10.632	95	220496	44.774	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049440.D
Acq On : 28 Jun 2022 06:36
Operator : SY/MD
Sample : N3482-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-38

Quant Time: Jun 28 07:16:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 05:14:22 2022
Response via : Initial Calibration

