

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
 Data File : VU049441.D
 Acq On : 28 Jun 2022 07:00
 Operator : SY/MD
 Sample : N3482-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-40

Quant Time: Jun 28 08:31:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 05:14:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	302399	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	520952	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	464983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	212204	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	201777	43.326	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.660%
35) Dibromofluoromethane	5.289	113	160636	44.074	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.140%
50) Toluene-d8	7.896	98	616442	47.420	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.840%
62) 4-Bromofluorobenzene	10.632	95	221888	44.558	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049441.D
Acq On : 28 Jun 2022 07:00
Operator : SY/MD
Sample : N3482-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-40

Quant Time: Jun 28 08:31:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 05:14:22 2022
Response via : Initial Calibration

