


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062722\
Data File : VU049395.D
Acq On : 27 Jun 2022 09:04
Operator : SY/MD
Sample : BFB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Title : SW846 8260
Last Update : Tue Jun 28 10:35:43 2022



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2878

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	2481	PASS
75	95	30	60	53.5	7005	PASS
95	95	100	100	100.0	13082	PASS
96	95	5	9	6.3	819	PASS
173	174	0.00	2	1.6	150	PASS
174	95	50	100	72.4	9477	PASS
175	174	5	9	7.5	711	PASS
176	174	95	101	96.9	9186	PASS
177	176	5	9	7.3	672	PASS