

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062818\
 Data File : VU025010.D
 Acq On : 28 Jun 2018 10:57
 Operator : MD/SY
 Sample : VU0628MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0628MBL01

Quant Time: Jun 29 05:29:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	219702	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	328946	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	305308	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	163310	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	165344	46.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.20%	
35) Dibromofluoromethane	4.89	113	130173	47.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.36%	
50) Toluene-d8	7.57	98	475245	47.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.46%	
62) 4-Bromofluorobenzene	10.31	95	177702	44.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062818\
Data File : VU025010.D
Acq On : 28 Jun 2018 10:57
Operator : MD/SY
Sample : VU0628MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0628MBL01

Quant Time: Jun 29 05:29:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

