

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062818\
 Data File : VU025021.D
 Acq On : 28 Jun 2018 15:29
 Operator : MD/SY
 Sample : J3707-22 20X
 Misc : 6.14g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB489-23-25-180622

Quant Time: Jun 29 06:48:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	229720	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	346115	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	320969	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	170099	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	173321	46.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.44%	
35) Dibromofluoromethane	4.89	113	134465	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
50) Toluene-d8	7.57	98	493974	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
62) 4-Bromofluorobenzene	10.31	95	185154	44.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
27) cis-1,2-Dichloroethene	4.23	96	20762	6.27	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062818\
Data File : VU025021.D
Acq On : 28 Jun 2018 15:29
Operator : MD/SY
Sample : J3707-22 20X
Misc : 6.14g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB489-23-25-180622

Quant Time: Jun 29 06:48:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

