

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062818\
 Data File : VU025025.D
 Acq On : 28 Jun 2018 17:06
 Operator : MD/SY
 Sample : J3707-26 20X
 Misc : 5.80g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB489-31-33-180622

Quant Time: Jul 06 18:10:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	222054	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	336771	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	307340	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	175870	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	165135	45.55	ug/l	0.00
Spiked Amount			Recovery	=	91.10%	
35) Dibromofluoromethane	4.89	113	131013	46.88	ug/l	0.00
Spiked Amount			Recovery	=	93.76%	
50) Toluene-d8	7.57	98	482738	47.35	ug/l	0.00
Spiked Amount			Recovery	=	94.70%	
62) 4-Bromofluorobenzene	10.31	95	186794	46.15	ug/l	0.00
Spiked Amount			Recovery	=	92.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
27) cis-1,2-Dichloroethene	4.23	96	122982	38.41	ug/l	88
39) Methylcyclohexane	6.42	83	8538	1.73	ug/l	95
44) Trichloroethene	6.19	130	471861	147.03	ug/l	96
78) n-propylbenzene	10.59	91	6701	0.42	ug/l	94
80) 1,3,5-Trimethylbenzene	10.78	105	17582	1.50	ug/l	98
84) 1,2,4-Trimethylbenzene	11.15	105	42918	3.60	ug/l	99
85) sec-Butylbenzene	11.33	105	8093	0.57	ug/l	99
86) p-Isopropyltoluene	11.48	119	6660	0.53	ug/l	89
88) 1,4-Dichlorobenzene	11.52	146	3830	0.60	ug/l #	25
89) n-Butylbenzene	11.89	91	20421	1.82	ug/l #	44
91) 1,2-Dichlorobenzene	11.89	146	6607	1.04	ug/l	86
95) Naphthalene	13.75	128	17210	1.99	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062818\
Data File : VU025025.D
Acq On : 28 Jun 2018 17:06
Operator : MD/SY
Sample : J3707-26 20X
Misc : 5.80µ/5.0mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB489-31-33-180622

Quant Time: Jul 06 18:10:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

