

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062818\
 Data File : VU025042.D
 Acq On : 29 Jun 2018 00:46
 Operator : MD/SY
 Sample : J3669-20ME
 Misc : 6.20g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB475-29-31-180620ME

Quant Time: Jun 29 08:07:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	225275	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	344309	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.10	117	333391	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	185631	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	171577	46.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.30%	
35) Dibromofluoromethane	4.89	113	131661	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
50) Toluene-d8	7.57	98	498570	47.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.66%	
62) 4-Bromofluorobenzene	10.32	95	204323	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.22	96	72024	22.17	ug/l	88
44) Trichloroethene	6.19	130	61075	18.61	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062818\
Data File : VU025042.D
Acq On : 29 Jun 2018 00:46
Operator : MD/SY
Sample : J3669-20ME
Misc : 6.20µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB475-29-31-180620ME

Quant Time: Jun 29 08:07:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

