

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049452.D
 Acq On : 28 Jun 2022 12:49
 Operator : SY/MD
 Sample : N3454-04
 Misc : 3.80G/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-15-1-2-GR

Quant Time: Jun 28 17:09:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	322954	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	559820	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.424	117	501166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	221476	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	214994	43.226	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.460%
35) Dibromofluoromethane	5.289	113	172471	44.035	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.080%
50) Toluene-d8	7.896	98	674624	48.292	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.580%
62) 4-Bromofluorobenzene	10.642	95	248008	46.345	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.700%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
18) Methyl Acetate	2.954	43	1890481	278.432	ug/l	93
25) 2-Butanone	4.729	43	76435	20.334	ug/l	97
39) Methylcyclohexane	6.755	83	42246	6.206	ug/l	93
40) Benzene	5.768	78	17904	1.044	ug/l	94
52) Toluene	7.970	92	80687	7.918	ug/l	99
59) 2-Hexanone	8.723	43	155103	25.918	ug/l	95
67) Ethyl Benzene	9.575	91	74988	3.987	ug/l	98
68) m/p-Xylenes	9.694	106	123372	17.727	ug/l	90
69) o-Xylene	10.105	106	60406	8.743	ug/l	90
73) Isopropylbenzene	10.488	105	16041	1.017	ug/l	98
78) n-propylbenzene	10.909	91	36873	1.923	ug/l	93
80) 1,3,5-Trimethylbenzene	11.092	105	72196	5.242	ug/l	95
84) 1,2,4-Trimethylbenzene	11.472	105	278980	20.617	ug/l	95
85) sec-Butylbenzene	11.649	105	9178	0.539	ug/l	100
86) p-Isopropyltoluene	11.796	119	60862	4.418	ug/l	94
89) n-Butylbenzene	12.208	91	43475	3.410	ug/l #	43
95) Naphthalene	14.086	128	246596	15.464	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049452.D
Acq On : 28 Jun 2022 12:49
Operator : SY/MD
Sample : N3454-04
Misc : 3.80G/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-15-1-2-GR

Quant Time: Jun 28 17:09:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

