

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049459.D
 Acq On : 28 Jun 2022 15:33
 Operator : SY/MD
 Sample : N3458-05 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP04-20220621

Quant Time: Jun 29 05:21:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	302414	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	522458	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	474583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	228075	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	198877	42.702	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.400%
35) Dibromofluoromethane	5.288	113	160949	44.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.060%
50) Toluene-d8	7.896	98	627035	48.095	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.200%
62) 4-Bromofluorobenzene	10.632	95	234634	46.982	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.960%
Target Compounds						
					Qvalue	
44) Trichloroethene	6.542	130	7286	1.915	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049459.D
Acq On : 28 Jun 2022 15:33
Operator : SY/MD
Sample : N3458-05 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP04-20220621

Quant Time: Jun 29 05:21:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

