

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049462.D
 Acq On : 28 Jun 2022 16:44
 Operator : SY/MD
 Sample : N3480-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 40815

Quant Time: Jun 28 17:18:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	308494	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	530679	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	485306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	233385	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	198896	41.864	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.720%
35) Dibromofluoromethane	5.289	113	164640	44.344	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.680%
50) Toluene-d8	7.896	98	640070	48.335	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.660%
62) 4-Bromofluorobenzene	10.632	95	234443	46.216	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.440%
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	4115939	1717.907	ug/l	99
25) 2-Butanone	4.710	43	30486	8.490	ug/l	# 89
52) Toluene	7.967	92	4229	0.438	ug/l	97
68) m/p-Xylenes	9.697	106	6739	1.000	ug/l	88
69) o-Xylene	10.102	106	1658	0.248	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049462.D
Acq On : 28 Jun 2022 16:44
Operator : SY/MD
Sample : N3480-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
40815

Quant Time: Jun 28 17:18:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

