

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049468.D
 Acq On : 28 Jun 2022 19:06
 Operator : SY/MD
 Sample : N3480-02DL 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 40815DL

Quant Time: Jun 28 19:31:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	310406	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	531149	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	473185	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	222675	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	204900	42.862	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.720%
35) Dibromofluoromethane	5.289	113	166321	44.757	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.520%
50) Toluene-d8	7.896	98	636923	48.055	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.100%
62) 4-Bromofluorobenzene	10.632	95	230131	45.326	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.660%
Target Compounds						
16) Acetone	2.630	43	206740	85.757	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049468.D
Acq On : 28 Jun 2022 19:06
Operator : SY/MD
Sample : N3480-02DL 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
40815DL

Quant Time: Jun 28 19:31:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

