

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049479.D
 Acq On : 28 Jun 2022 23:49
 Operator : SY/MD
 Sample : N3386-07DL 100X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE108D2-20220614DL

Quant Time: Jun 29 05:24:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	303604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	522442	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	460020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	204191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	201959	43.193	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.380%
35) Dibromofluoromethane	5.288	113	162456	44.446	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.900%
50) Toluene-d8	7.896	98	622668	47.762	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.520%
62) 4-Bromofluorobenzene	10.632	95	219398	43.932	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.860%
Target Compounds						
44) Trichloroethene	6.542	130	22801	5.994	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049479.D
Acq On : 28 Jun 2022 23:49
Operator : SY/MD
Sample : N3386-07DL 100X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE108D2-20220614DL

Quant Time: Jun 29 05:24:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

