

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049481.D
 Acq On : 29 Jun 2022 00:37
 Operator : SY/MD
 Sample : N3457-07 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE115D2-20220620

Quant Time: Jun 29 05:24:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	313627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	538036	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	478979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.813	152	222623	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	208397	43.146	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.300%
35) Dibromofluoromethane	5.289	113	168051	44.644	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.280%
50) Toluene-d8	7.896	98	639898	47.661	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.320%
62) 4-Bromofluorobenzene	10.633	95	231916	45.093	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.180%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.578	101	14280	4.104	ug/l	97
12) 1,1-Dichloroethene	2.578	96	6746	1.980	ug/l	96
44) Trichloroethene	6.543	130	308126	78.656	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049481.D
Acq On : 29 Jun 2022 00:37
Operator : SY/MD
Sample : N3457-07 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE115D2-20220620

Quant Time: Jun 29 05:24:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

