

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049484.D
 Acq On : 29 Jun 2022 01:47
 Operator : SY/MD
 Sample : N3457-13DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE103D3-20220620DL

Quant Time: Jun 29 05:25:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	307233	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	533404	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	470365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.809	152	217315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	205267	43.382	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.760%
35) Dibromofluoromethane	5.289	113	167101	44.777	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.560%
50) Toluene-d8	7.896	98	633709	47.610	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.220%
62) 4-Bromofluorobenzene	10.632	95	228154	44.747	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.500%
Target Compounds						
					Qvalue	
44) Trichloroethene	6.543	130	67563	17.397	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049484.D
Acq On : 29 Jun 2022 01:47
Operator : SY/MD
Sample : N3457-13DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE103D3-20220620DL

Quant Time: Jun 29 05:25:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

