

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049485.D
 Acq On : 29 Jun 2022 02:11
 Operator : SY/MD
 Sample : N3457-09 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE132D6-20220620

Quant Time: Jun 29 05:25:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	297754	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	527332	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	470806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	216933	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	203665	44.414	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.820%
35) Dibromofluoromethane	5.288	113	163830	44.406	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.820%
50) Toluene-d8	7.899	98	623085	47.351	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.700%
62) 4-Bromofluorobenzene	10.632	95	228028	45.237	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.480%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.581	101	3904	1.182	ug/l	94
12) 1,1-Dichloroethene	2.578	96	1260	0.390	ug/l	96
44) Trichloroethene	6.542	130	254509	66.288	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049485.D
Acq On : 29 Jun 2022 02:11
Operator : SY/MD
Sample : N3457-09 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE132D6-20220620

Quant Time: Jun 29 05:25:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

