

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049486.D
 Acq On : 29 Jun 2022 02:34
 Operator : SY/MD
 Sample : N3457-10 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE132D7-20220620

Quant Time: Jun 29 05:25:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	309923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	536051	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	477587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	220916	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	205428	43.039	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	86.080%		
35) Dibromofluoromethane	5.288	113	165141	44.034	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.060%		
50) Toluene-d8	7.896	98	639392	47.800	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	95.600%		
62) 4-Bromofluorobenzene	10.632	95	229526	44.793	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.580%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.578	101	2836	0.825	ug/l	91
44) Trichloroethene	6.542	130	193563	49.595	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049486.D
Acq On : 29 Jun 2022 02:34
Operator : SY/MD
Sample : N3457-10 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE132D7-20220620

Quant Time: Jun 29 05:25:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

