

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049491.D
 Acq On : 29 Jun 2022 04:32
 Operator : SY/MD
 Sample : N3457-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW179D-20220617

Quant Time: Jun 29 05:26:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	293868	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	517577	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	470335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	218444	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	202080	44.651	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.300%
35) Dibromofluoromethane	5.289	113	161535	44.609	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.220%
50) Toluene-d8	7.896	98	617664	47.823	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.640%
62) 4-Bromofluorobenzene	10.632	95	224766	45.430	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.860%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.581	96	818	0.256	ug/l	90
19) Methyl tert-butyl Ether	3.363	73	7972	0.672	ug/l	93
27) cis-1,2-Dichloroethene	4.668	96	77229	19.427	ug/l	99
44) Trichloroethene	6.543	130	314408	83.433	ug/l	87
64) Tetrachloroethene	8.552	164	33850	11.221	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049491.D
Acq On : 29 Jun 2022 04:32
Operator : SY/MD
Sample : N3457-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW179D-20220617

Quant Time: Jun 29 05:26:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

