

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
 Data File : VU049498.D
 Acq On : 29 Jun 2022 07:16
 Operator : SY/MD
 Sample : N3457-20
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW178S-20220621

Quant Time: Jun 29 08:03:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	290789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	506167	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	461159	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	217590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	197948	44.201	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.400%
35) Dibromofluoromethane	5.288	113	157623	44.510	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.020%
50) Toluene-d8	7.899	98	609443	48.251	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.500%
62) 4-Bromofluorobenzene	10.632	95	222882	46.065	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.120%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	3.366	73	14393	1.225	ug/l	96
44) Trichloroethene	6.542	130	3590	0.974	ug/l	94
64) Tetrachloroethene	8.552	164	632	0.214	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062822\
Data File : VU049498.D
Acq On : 29 Jun 2022 07:16
Operator : SY/MD
Sample : N3457-20
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW178S-20220621

Quant Time: Jun 29 08:03:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

