

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062823\
 Data File : VU054685.D
 Acq On : 28 Jun 2023 17:30
 Operator : MD/SY
 Sample : 03457-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0C51

Quant Time: Jun 29 04:37:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 23 23:55:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	303900	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	294582	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	136991	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	94902	4.907	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.200%
7) Chloroethane-d5	1.911	69	82147	4.536	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.800%
11) 1,1-Dichloroethene-d2	2.563	65	38155	4.417	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.400%
20) 2-Butanone-d5	4.634	46	309278	56.876	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.760%
24) Chloroform-d	5.058	84	172009	4.255	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
26) 1,2-Dichloroethane-d4	5.698	65	93947	4.767	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.400%
32) Benzene-d6	5.724	84	335458	4.097	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.000%
36) 1,2-Dichloropropane-d6	6.685	67	120671	4.413	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.200%
41) Toluene-d8	7.894	98	283396	3.900	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.000%
43) trans-1,3-Dichloroprop...	8.177	79	43598	4.584	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.600%
46) 2-Hexanone-d5	8.631	63	205204	51.432	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.860%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	96793	4.746	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	102928	4.455	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062823\
Data File : VU054685.D
Acq On : 28 Jun 2023 17:30
Operator : MD/SY
Sample : 03457-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0C51

Quant Time: Jun 29 04:37:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 23 23:55:46 2023
Response via : Initial Calibration

