

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062919\
 Data File : VU033069.D
 Acq On : 28 Jun 2019 13:29
 Operator : JC/SP
 Sample : VU0629WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK08

Quant Time: Jun 29 00:38:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 29 00:32:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	95459	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	95321	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	48686	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	28244	4.45	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.00%
7) Chloroethane-d5	1.68	69	24378	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.80%
11) 1,1-Dichloroethene-d2	2.27	63	46043	3.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.80%
20) 2-Butanone-d5	4.20	46	93854	48.65	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.30%
24) Chloroform-d	4.64	84	59847	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
26) 1,2-Dichloroethane-d4	5.30	65	35677	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.60%
32) Benzene-d6	5.33	84	111540	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
36) 1,2-Dichloropropane-d6	6.32	67	36541	4.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.20%
41) Toluene-d8	7.56	98	103036	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
43) trans-1,3-Dichloropropene-	7.85	79	14998	4.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.40%
46) 2-Hexanone-d5	8.31	63	63136	47.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.64%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	30566	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.20%
64) 1,2-Dichlorobenzene-d4	11.85	152	42332	4.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062919\
Data File : VU033069.D
Acq On : 28 Jun 2019 13:29
Operator : JC/SP
Sample : VU0629WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK08

Quant Time: Jun 29 00:38:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 29 00:32:07 2019
Response via : Initial Calibration

