

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062919\
 Data File : VU033085.D
 Acq On : 28 Jun 2019 21:07
 Operator : JC/SP
 Sample : K3597-09
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAM66

Quant Time: Jun 29 01:26:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 29 00:59:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	82137	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	86218	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	40860	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	35683	6.53	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	130.60%#
7) Chloroethane-d5	1.68	69	27291	6.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	122.20%
11) 1,1-Dichloroethene-d2	2.27	63	53821	4.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.80%
20) 2-Butanone-d5	4.19	46	88785	53.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.98%
24) Chloroform-d	4.64	84	61138	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.80%
26) 1,2-Dichloroethane-d4	5.30	65	33750	5.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.00%
32) Benzene-d6	5.33	84	113993	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.80%
36) 1,2-Dichloropropane-d6	6.32	67	35979	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	7.56	98	99119	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
43) trans-1,3-Dichloropropene-	7.85	79	14312	4.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.40%
46) 2-Hexanone-d5	8.31	63	58562	48.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.04%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	29543	4.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.60%
64) 1,2-Dichlorobenzene-d4	11.85	152	39011	5.05	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.00%

Target Compounds					Ovalue
13) Acetone	2.32	43	9980	7.135 ug/L	97
21) 2-Butanone	4.29	43	5145	2.613 ug/L	96
47) Tetrachloroethene	8.22	164	237173	42.355 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062919\
Data File : VU033085.D
Acq On : 28 Jun 2019 21:07
Operator : JC/SP
Sample : K3597-09
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAM66

Quant Time: Jun 29 01:26:03 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 29 00:59:26 2019
Response via : Initial Calibration

