

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062920\
 Data File : VU039191.D
 Acq On : 29 Jun 2020 15:14
 Operator : SY/MD
 Sample : L3139-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jun 30 02:01:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 30 01:35:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	263091	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	266058	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	116634	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	94291	3.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.20%
7) Chloroethane-d5	1.93	69	78192	4.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.80%
11) 1,1-Dichloroethene-d2	2.59	63	135656	3.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.20%
20) 2-Butanone-d5	4.68	46	324296	43.97	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.94%
24) Chloroform-d	5.10	84	173658	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.00%
26) 1,2-Dichloroethane-d4	5.74	65	110551	4.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.00%
32) Benzene-d6	5.76	84	326972	3.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.80%
36) 1,2-Dichloropropane-d6	6.72	67	108618	4.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.20%
41) Toluene-d8	7.92	98	258500	3.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.00%
43) trans-1,3-Dichloropropene-	8.20	79	46710	3.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.80%
46) 2-Hexanone-d5	8.66	63	236563	42.19	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.38%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	92550	3.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	75.80%
64) 1,2-Dichlorobenzene-d4	12.20	152	106033	4.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.60%

Target Compounds					Ovalue
16) Methylene chloride	3.08	84	3556	0.153 ug/L	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062920\
Data File : VU039191.D
Acq On : 29 Jun 2020 15:14
Operator : SY/MD
Sample : L3139-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jun 30 02:01:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 30 01:35:59 2020
Response via : Initial Calibration

